

Oscilloscope probe

Model: 011040

SPECIFICATION

Attenuation Ratio	1:1
Bandwidth	DC to 40MHz
Rise Time	8.75ns
Input Resistance	1M(Oscilloscope Input)
Input Capacitance	46pF plus oscilloscope capacitance
Working Voltage	300V DC incl. Peak AC Derating with frequency
Safety	conformed IEC-1010
Cable Length	0.9Meter

MAINTENANCE

Before dismantling any part of the probe, make sure it is disconnected from any voltage source. The probe body can be detached from the cable assembly by unplugging the push fit BNC connector from the probe body. This permits replacement of the cable or body assembly should either part become damaged.



北京海洋兴业科技股份有限公司 (证券代码: 839145)

北京市西三旗东黄平路19号龙旗广场4号楼 (E座) 906室

邮编: 100096

电话: 010-62176775 62178811 62176785

传真: 010-62176619

企业QQ: 800057747 维修QQ: 508005118

邮箱: market@oitek.com.cn

企业官网: www.hyxyyq.com

购线网: www.gooxian.com



扫描二维码关注我们
重庆微信公众号: 海洋仪器